



Dated 07.05.2018

**CORRIGENDUM**

With reference to tender no. AIIMS/Pat/Tender/2018/Gastro./F9117 regarding Supply & Installation of Portable Ultrasound Machine with Provision for Doppler, a pre-bid meeting was held on 21/04/2018 at Medical College Building. The issues raised by different companies were discussed and the committee recommended the following corrigendum:

Point no as in original specs	Specification as per original tender	Revised specifications Should be read as
3.	The unit should be compact, lightweight and portable.	The unit should be compact, lightweight (<6. Kg) and portable.
10	(1) Convex electronic phased array transducer: 2-6 MHz for abdominal imaging. (2) Linear transducer: 5-12MHz MHz for vascular and small part imaging.	(1) Convex electronic phased array transducer: 2-6 $\pm$ 1 MHz for abdominal imaging. (2) Linear transducer: 5-14 $\pm$ 2 MHz for vascular and small part imaging.
13.	Provision for attaching an external keyboard and mouse should be present.	This is optional
14	The System must have integrated high – resolution TFT/LCD/Single monitor of 15 Inches or more.	The System must have integrated high – resolution LCD/Single monitor of 11.5 Inches or more.
21	In built battery backup should be at least one hour or more.	In built battery backup should be at least 45 minutes or more.
25	Comprehensive Warranty for 5 years followed by 5 years of free AMC	Comprehensive Warranty for 5 years followed by 5 years of CAMC

Bid Submission date extended 19.05.2018.

S/D  
Faculty in charge  
Procurement cell  
AIIMS Patna