



07/08/2018

Corrigendum /addendum

With reference to tender number AIIMS/pat/tender/2018/Dentistry/11557 for supply and installation of Supply and Installation of Dental chair unit for Department of dentistry a pre bid meeting was held on 28/07/2018 at Medical College Building, administrative block, AIIMS Patna.

After discussion with their representatives committee recommends Following addition/ deletion / amendments suggested.

Patient Chair: 1st point:

Which is regarding the use of servometer should be read as Imported Motor which would offer motion sequence while working in seated or standing position.

Patient Chair: 16th point:

The pump cover should be ABS should be read as complete chair plastic part as ABS quality with fire retardant coating or molded of a similar fire retardant material.

S/D
Faculty in charge
Procurement cell
AIIMS Patna