



Dated: 24/10/2018

## CORRIGENDUM

With reference to tender number AIIMS/Pat/Tender/ICP/pediatric/2036 for supply and installation of ICP Monitoring a pre bid meeting was held on 12/10/2018 at Medical College Building, administrative block, AIIMS Patna. The issues raised by two companies were discussed and the committee recommended the following corrigendum:

After discussion regarding above- mentioned request from potential bidder following amendments are made in specification "Notice Inviting Tender for Supply and installation of ICP MONITORING for Department of Pediatrics of AIIMS PATNA"

Point no as in Original specification	Specification as per original tender	Revised specifications Should be read as
<b>It should include following accessories:</b>		
6	Rechargeable 3-hour battery operation for patient transport	Rechargeable at least 1-hour battery backup operation for patient transport

SD/  
Faculty In-charge  
Procurement Cell  
AIIMS, Patna