



Dated 01.11.2018

**CORRIGENDUM**

With reference to tender no: **AIIMS/pat/tender/2018/PMR/NCV LAB/2018/2703** dated 05.10.2018 as per Pre-bid meeting held at committee room on 12.10.2018 at 12:00 PM

The issue raised by the firms were discussed and the committee recommended for the following corrigendum

1. As per tender specification Minimum 4 channel..... **Should be read as** Minimum 4 -8 channel system with optical isolation with Ethernet/USB connection for connecting to either to desktop system or laptop system for portable use.
2. As per tender specification the electrical stimulator..... **Should be read as** the electrical stimulator should have controls for stimulus delivery, intensity, store, reverse polarity button and two or more programmable buttons for inbuilt functions for controlling the system from the stimulator itself.
3. As per tender specification Constant current stimulator ..... **Should be read as** Constant current stimulator with current variable from 0 to 100mA with increments of 0.5mA and pulse duration to be varied from 50micro sec or less – 1000micro secs with 50micro sec increments or less.  
**Tech specs under “Amplifier should have “: should be read as** Common mode input impedance > 1000Mohm, Gain to be varied from 10 nV/div to 100mV/div or higher.
4. **Tech specs under “.EMG Should have”:** **read as** Should have facility of Anatomy ..... **should be read as** Should have facility of Anatomy view feature displays the EMG muscle scoring data on a 2D/ 3D colour-coded anatomical model, preferably 3D as it will be also helpful for teaching training research and academic purpose.  
Auto finding composer which can insert predefined sentences and print the report in MS word and other standard format.  
Facility to wave form and numerical data as per user requirement preferably with anatomic view for teaching training research and academic purpose.
5. **Tech specs under “EP should have”:** 15” VEP monitor with.....**should be read as** 20” or more VEP monitor with checker board patterns, half, quanta size.
6. **The system should be supplied with: mentioned all items under this heading should be read as mention below:**  
Latest processor -i5 -i7 processor or superior  
6 to 8 GB RAM or superior  
1 TB hard disc or superior  
20 inch or more LED monitor  
LaserJet printer  
System should have appropriate UPS backup.  
Portable trolley with caster wheel.
7. TENDER SPECIFICATION  
US/FDA/BIS/EUROPEAN CE

Tender submission date till 16.11.2018 time 12:00 and received bids will open on same day time 12.30 PM.

Faculty in charge  
Procurement cell