



Corrigendum

Dated: 23-02-2018

With reference to tender AIIMS/Pat/Tender/Forensic/Instrument/2017- 2843 dated 12/01/2018 pre-bid meeting was held on 19-01-2018 at 12:00 PM in conference hall main building AIIMS Patna. The following changes in specification should be published as corrigendum:

Following addition/ deletion / amendments suggested

1. In clause '6' of eligibility criteria on page '6' Average Annual turnover during last three years required should be reduced to Rs 75 lakh from Rs 1 crore.
2. In clause 11a and 25 of the General term condition of the contract under the heading of Installation and delivery period, following lines should be added: - In case of items to be supplied from Foreign Country 90 days from the date of issuance of supply order/ opening of error less LC.
3. In clause 27 (ii) a of General term condition of contract in place of "75% of payment of the contract price shall be paid 60 days after presentation of shipment documents." Should be replaced by "75% of payment of the contract price shall be paid immediately against submission of shipping documents like airway bill or bill of loading".
4. In clause 22 of check list of terms and conditions certificate to the effect of two supply order should be amended as one.
5. In clause 21 h of General term condition of the contract the word "Deputy Director" should be replaced by deputy director/MS.
6. In annexure "XI" the specification should be as follows: -

Modified/replaced specification for Portable scale hanger with double dial hanging scale (Qty 2) (after pre-bid meeting held on 19/01/2018)

- Should have portable scale hanger with steel base.
- Should have stationary pole secured to base.
- Length 72 in. /182 cm.
- Should have 3 in. /7 cm. swivel casters, (2) locking, (2) non locking
- Should be able to hang **Double Dial Hanging Autopsy Scale.**

The Double Dial Hanging Autopsy Scale should have the following features:

- Made of heavy chrome scale body and pan suspension.
- Should be accurate to 0.1% of full scale and with Double 13 inch dial for reading from either side.
- Should come with bow, ring and pan and with following technical specification:

Diameter (Scale): 15.875 in. / 40 cm

Depth (Scale): 3.4875 in. /12 cm

Diameter (Pan): 14.5 in. / 36 cm

Depth (Pan): 5 in. /12 cm

- Capacity by graduation: 15 Kg x 25 g with pan hang.

Should have European/USA, CE/FDA certificate.

2. Tender Submission last date extended till dated 07.03.2018

Faculty in charge
Procurement cell
AIIMS Patna