



Dated: 27/08/2018

**Corrigendum /Addendum**

With reference to tender number AIIMS/Pat/Tender/Admn/PFB/3852 for supply and installation of Pediatric Flexible Bronchoscope a pre bid meeting was held on 08/08/2018 at Medical College Building, administrative block, AIIMS Patna.

After discussion regarding above- mentioned request from potential bidder following amendments are made in specification "Notice Inviting Tender for Supply and installation of Pediatric Flexible Bronchoscope for Paediatrics of AIIMS PATNA"

Point no as in original specification	Specification as per original tender	Revised specifications Should be read as
A (3)	Should be compatible with Analog, HD-SDI and DVI output & 16:9 & 16:10 output for a HDTV monitor should be available	Should be compatible with Analog, HD-SDI or DVI output & 16:9 & 16:10 output for a HDTV monitor should be available
A (5)	Compact, lightweight (10-11 kg) and ergonomically designed	Compact, lightweight (10-14 kg) and ergonomically designed
A (6)	Should have band imaging facility like NBI, FICE or Full	Should have band imaging facility like NBI/optical enhancement i-scan
A (7)	i-scan capacity for compatibility with NBI, Fice or i-scan Video scopes.	Line delete
A (10)	Equipped with memory backup for settings & Lithium battery	Equipped with memory backup
A (15)	Equipped with memory backup for settings & Lithium battery.	Line delete
B. (1)	Equipped with Band Imaging capability with high intensity Xenon Light source (300W) or LED light source.	Equipped with Band Imaging capability with high intensity Xenon Light source (300W).
B. (6)	Should have pre freeze function for image stabilization Light source (Xenon short arc Ozone free 300 Watt lamp / LED):-	Should have pre freeze function for image stabilization Light source (Xenon short arc Ozone free 300 Watt lamp) with at least 500hr guarantee & 2 Xenon bulbs extra.
C. (4)	Insertion Outer diameter should be 2.8 mm or less	Insertion Outer diameter should be 3.8 mm or less
C. (5)	Working Channel diameter should be 1.2 mm or less.	Working Channel diameter should be 1.2 mm or more
C. (14)	High Resolution Medical Grade LCD Monitor: - 26 inch full HD LCD monitor with high resolution.	High Resolution Medical Grade LCD Monitor: - 24 inch full HD LCD monitor with high resolution.
C. (16) a.	Mini Biopsy Forceps oval rat forceps: 2	Mini Biopsy Forceps oval rat forceps: 10
C. (16) b.	Mini Cytology Brushes: 2	Mini Cytology Brushes: 10
C. (16) c.	Three-Prong Grasping Forceps opening 10mm: 2	Three-Prong Grasping Forceps opening 10mm: 10
<b>Additional Points</b>		
C. (16) f.	Trolley (High quality)	
C. (16) g.	High definition display	
C. (16) h.	Image Capturing software	
C. (16) i.	Data archiving and recording system	
C. (16) j.	Computer System with at least core-i3 OS Windows-7, 4 GB RAM	
C. (16) k.	Good Quality portable suction machine	



## ALL INDIA INSTITUTE OF MEDICAL SCIENCES PATNA



C. (16) l.	Colour laser Printer
C. (16) m.	Online UPS (2 KVA)
C. (19)	The whole System should be 5 yrs. warranty & after that 5 year CMC

The date and time for submission and opening of tender document extended from 28.08.2018 at 12:00 hrs to 04.09.2018 12:00 hrs.

SD/  
Faculty In-charge  
Procurement Cell  
AIIMS, Patna