



Dated 26.12.2018

CORRIGENDUM

With reference to tender No AIIMS/Pat/Tender/admn/PFB/3852 dated 29/11/2018 for Supply and Installation of Pediatric Flexible Bronchoscope for Department of Pediatric at AIIMS Patna, pre-bid committee has proposed following amendment as mentioned below:

Tender document Pg. no	As tender document	Should be read as
A (5)	Compact, lightweight (10-11 Kg) and ergonomically designed	Compact, lightweight and ergonomically designed.
B.	Light Source (Xenon short arc ozone free 300 watt lamp/Led)	Light Source (Xenon short arc ozone free 300 watt lamp.
B. (4)	Emergency halogen light for backup.	Emergency light for back up:-
B.6 Second line	10 emergency halogen light for backup: - Equipped with automatic light adjustment forced cooling, regulated air feeding pump and fan with low noise.	10 emergency halogen light for backup: - Equipped with automatic light adjustment forced cooling, regulated air feeding pump and fan with low noise.
C. (8)	Depth of field should be 2-50 mm or better.	Depth of field should be 3-50 mm or better.
C. (9)	Left rotation and right rotation 120/120 degree.	Delete line.
C. (16) a	Mini Biopsy forceps oval rat forceps:10	Mini Biopsy forceps oval rat forceps (standard make): 10
C. 18	The Equipment should be USFDA and European CE approved	The Equipment should be USFDA/European CE approved.
Addendum		
Schedule of requirement	One as per Annexure XI	Should be read as Two
EMD in Rs.	60000.00	120000.00

Tender submission date will be extended from 22.12.2018 to 04.01.2019 till 12:00 PM and submitted bids will be opened same day time 12:30.

Sanyu
SD/ 26/12/18

Faculty In-charge
Procurement Cell
AIIMS, Patna