



Inquiry No. AIIMS/Pat/Store/LP/Pulmonary Equipments/2014-15/016

Date: - 15/12/2014

Invitation of quotation for Pulmonary Equipments

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Patna for supply of Pulmonary Equipments to the Department of Pulmonary Medicine as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 26.12.2014 12.00 hrs. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR THE PULMONARY EQUIPMENTS AGAINST

INQUIRY NO. AIIMS/Pat/Store/LP/Pulmonary Equipments/2014-15/016” DUE ON 26.12.2014 12.00 HRS”

1. **Terms & Conditions:**

- a) Envelope should be super-scribed “QUOTATION FOR THE PULMONARY EQUIPMENTS AGAINST THE INQUIRY NO. AIIMS/Pat/Store/LP/Pulmonary Equipments/2014-15/016” DUE ON 26.12.2014 12.00 HRS. . **Quotations need to be submitted by speed post/registered post or may be dropped in the tender box placed in Administrative Office, Medical College building after obtaining the acknowledgement for the same in the office of Administrative Officer, AIIMS Patna.**
The quotations received after this deadline shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
- b) **Unsealed quotation will be rejected.**
- c) Quotations must be in the enclosed prescribed Performa and forwarding letter on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, in case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- d) Rates must be quoted as per the format specified taxes extra if any must be written separately. The rates must be quoted in figures as well as in words.
- e) In general no overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected. However, except rate all cuttings and over writings must be signed by the authorized person of the firm.
- f) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- g) Ceiling amount of this quotation will be 1 Lac.

- h) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
L1 will be decided for the overall value of quotation and not item wise
- i) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- j) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Bihar/ Central Government.
 - The firm shall have valid VAT/ Sales Tax No. and IT PAN.
 - **Certificate of non-inclusion in the black list as per given format attached in annexure "2" need to be provided on Rs. 100/- stamp paper duly notarized.**
 - Market standing of minimum 3 Years.
- k) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- l) The supplier may be asked to submit a sample of the product, which will be evaluated by a technical evaluation committee. Price bids of only those firms will be opened which qualifies technically as per the recommendation of the committee.
The expenditure incurred for demonstrating the items will be borne by the supplier.
- m) **Delivery Period** – 15 days from award of work.
- n) **Liquidated Damage:** -In the event of the Seller's failure supply the stores/goods are conduct trails, installation of equipment, training, etc. as specified in this inquiry, the Buyer may at his discretion, withhold any payment until the completion of the contract. The Buyer may also deduct from the Seller as agreed, liquidated damages to the sum of 0.5% of the contract price of the delayed/undelivered, stores/services mentioned above for every week of delay or part their of a week, subject to the maximum value of the Liquidated Damages being not higher that 10% of the value of delayed stores.
- o) **Guarantee/Warranty Terms:** The equipment's/Instruments/Materials supplied by result of this Quotation/supply order shall be of the best quality and workmanship and new in all respects and shall be strictly in accordance with the specifications and particulars contained /mentioned against each item in this document/ supply order. The seller guarantees that the said good equipment's / Instruments would continue to confirm to the description and quality a foresaid for a period of 12 months from the date of delivery/installation of the said equipment's/Instruments/materials to the buyer.
- p) **Payment Terms:** 100% after delivery and submission of following documents by vendors to Finance Accounts Section -
- a) Contingent Bill in triplicate along with supporting vouchers.
 - b) Store Receipt Certificate issued by stores.
 - c) Installation/Inspection Report issued by the concerned department if applicable.
- q) **Disputes:** - In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Patna with regards to the

interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Patna whose decision will be final and binding upon the contractor.

- r) AIIMS, Patna reserves the right to increase or decrease amount of work. Decision of Quantity of Equipment and Instruments in the AIIMS, Patna will be final in this regard.
- s) AIIMS, Patna reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, PATNA will be final in this regard.

Faculty-in-charge AO

Encl.: Annexure 1 (Format of Price Bid) Annexure 2 (Declaration Format) Annexure 3 (Specification)

PRICE BID FORM

To,
The Faculty-in-charge AO,
AIIMS Patna.

Dear Sir,

1. I/We submit the quotation for Enquiry No. **"QUOTATION FOR THE PULMONARY EQUIPMENTS AGAINST THE INQUIRY NO. AIIMS/Pat/Store/LP/Pulmonary Equipments/2014-15/016" DUE ON 26.12.2014 12.00 HRS. .** for "Supply of Pulmonary Equipments in the department of Pulmonary Medicine" at AIIMS Patna.
2. I/We thoroughly examined and understood terms & conditions of contract given in the enquiry document.
3. I/We hereby offer to supply at the following rates. I/ We undertake that I/ We are not entitled to claim any enhancement of rates on any account during the validity of rate.

Sl. No.	Product with Description	Make & Model No.	Unit	Quantity	Price per unit	Taxes (Including all @ %.)	Price per Unit (inclusive of all taxes)	Total Amount	Total Amount (In Words)
(a)	(b)	(c)	(d)	(e)	(f)	(g)	(h = f + g)	(I =e x h)	(j)
1.	BiPAP Machine (Specification attached in annexure "3")		Nos.	1					
2.	Nebulizer(Specification attached in annexure "3")		Nos.	1					
3.	Pulse Oximeter (Specification attached in annexure "3")		Nos.	2					

(Signature of Authorized Person)

(Name)

(Designation)

Name of Firm/Company/Agency

Place- _____

Date- _____

Contact Detail-_____

ANNEXURE-“2”

DECLARATION

Date.....

To,

The Faculty-in-charge AO,
AIIMS Patna.

Dear Sir,

Ref.: Inquiry No.

I / We hereby confirm that our firm has not been banned or blacklisted by any government organization/Financial institution/Court /Public sector Unit /Central Government and no police/ Vigilance enquiry/criminal case is pending against us.

Signature of Authorized Person.....

Name.....

Designation.....

Seal

Place:

Date:

Specification

1. Pulse Oximetre

- Dual Colour OLED Display
- 6 Display Modes
- Suitable for Both Adults and Pediatric
- Low Power Consumption
- Auto power off

2. Nebulizer

Type	Compressor Nebulizer
Electrical	230V, 50Hz model; 230V, 60Hz model
Nebulizer rate	0.4ml/min (without cap)
Particle size	MMD approximately 5 micrometers (based on EN13544-1)
Medication cup capacity	7ml (cc) max.

Appropriate medication quantities	2 - 7ml (cc)
Operating Temperature / Humidity	plus 10deg C to +40deg C, 30% to 85% RH
Storage Temperature / Humidity	minus 20deg C to +60deg C, 10% to 95% RH

Features

- Usable to both adults and child
- MMD approximately 5 micrometer
- Nebulizer rate : 0.4 milliliter per minute (without cap)
- Contains: Compressor, nebulizer kit, air tubes, 5 pieces replacement filters, mouth piece, adult mask, child mask, instruction manual and carrying bag
- 1 year manufacturer's warranty from the date of purchase

3. BIPAP

- Non Invasive ventilator with following modes:
CPAP,S,S/T, PC and T
- Should have IPAP:4-25 cmH₂o
EPAP:4-25 cmH₂o
CPAP:4-20 CmH₂o
- Breath Rate: Should be able to set up to 25 BPM.
- Should be provided with integrated heated humidifier.
- Should have a carry bag.
- Compliance data should be stored on external memory card which can be accessed through non proprietary card readers.
- **Should provide flow based pressure relief at transition from exhalation to inhalation, transition from inhalation to exhalation and during exhalation**

- **System should be able to compensate for variable resistance related to different mask.**
- System should have advanced technology to prevent rain out.
- **Humidifier chamber should be isolated from the inner parts of the device to prevent damage.**
- Should be USA FDA approved.
- Should have warranty of 2 years
- **Manufacture should have wide national network and office with complete service, repair and application backup capability for the quoted equipment.**
- Device should ensure the delivery of targeted tidal volume for non invasive ventilation patients (up to 1500 ml)
- Should have following alarms:
Patient disconnect , Apnea, Low minute ventilation , Low Tidal Volume
- Automatic triggering and cycling to ensure best patient to machine synchronization.
- Should be provide with latex free full face reusable mask Qty-2
- Should be provided with hose pipe Qty.-2
- Facility for downloading compliance data from the equipment at the hospital has to be provided.
- Should have facility to monitor parameters like Respiratory rate ,Exhaled tidal volume , Leak , Minute ventilation.