



Dated: 06/12/2017

CORRIGENDUM

With reference to tender no: AIIMS/Pat/Tender/2017/GenSurgery/CUSA/396 dated 14/11/2017, regarding Supply and Installation of Cavitron Ultrasonic Surgical Aspirator (CUSA) for Department of General Surgery a pre-bid meeting was held on 22/11/2017 in committee hall, medical college building, AIIMS Patna

The issue raised by the firms were discussed and the committee recommends for the following corrigendum to be issued:-

1. **In Technical Specification point 7 (a)-** The frequency range to be in the range of 23-36 KHz
2. **In Technical Specification point 7 (c)-** The frequency range to be in the range of 23-36 KHz
3. **In Technical Specification point 33 –** All items should be from OEM except the UPS, ESU and sterile container for the handpieces
4. The last date of submission of tender documents may be extended up to 20-12-2017, 14.00 hrs and to open on 20/12/2017 at 15:30 hrs.

sd/

Faculty In-charge

Procurement Cell

AIIMS, Patna