



Dated: 06/12/2017

CORRIGENDUM

With reference to tender no: AIIMS/Pat/Tender/PedSur/InfusionPump/5606 dated 14/11/2017, regarding Supply & Installation of Neonatal Infusion Pump a pre-bid meeting was held on 21/11/2017 in committee hall, medical college building, AIIMS Patna

The issue raised by the firms were discussed and the committee recommends for the following corrigendum to be issued:-

- 1) Fourth Point- Which is flow rate programmable should be changed as **minimum infusion rate 0.01 ml/Hr.**
- 2) Fifth Point-"Display of drug name" **The committee recommends omission of this point.**
- 3) Seventh Point- "Rechargeable battery with upto six hours back up". **The committee recommends Rechargeable battery back up should be in the range of 06-10 Hrs.**
- 4) One point addition is recommended by the committee- **There should be stacking facility for pumps.**
- 5) **USFDA/CE approved certification required.**
- 6) **The heading should be changed from "Neonatal Infusion Pump" to "Syringe Infusion Pump".**
7. The last date of submission of tender documents may be extended up to 22-12-2017, 14.00 hrs and to open on 22/12/2017 at 15:30 hrs.

sd/

Faculty In-charge

Procurement Cell

AIIMS, Patna