



Date: 30.10.2018

**CORRIGENDUM**

With reference to tender no: AIIMS/Pat/Tender/OPH/18-19/F-12349 tender for Supply and Installation Slit Lamp Biomicroscope for Department of Ophthalmology at AIIMS Patna, a pre-bid meeting was held on 15/05/2018 in committee hall, medical college building, AIIMS Patna.

The issue raised by the firm where discussed and the committee recommended for the following Corrigendum/revised specification replaced with tender specification:-

**REVISED TECHNICAL SPECIFICATIONS FOR SLIT LAMP BIOMICROSCOPE**

1. Assembly of Haag- Streit type
2. Eye piece magnification 10x or 12.5x
3. Overall magnification range 6.0x to 40x ( $\pm 0.5$  in 5 steps)
4. Slit rotation 0-180°
5. Slit width range 0-8mm or more.
6. Slit length range 1-8mm or more.
7. Slit decentration facility  $\pm 4^\circ$  or more.
8. Built in filters; cobalt blue, red free, heat absorbing.
9. Chin rest adjustment facility.
10. Light source; LED, illumination ranges 10,000 Lux to 4,50,000 Lux.
11. Illumination control adjustable.
12. Patient fixation light.
13. Goldmann type Applanation tonometer appropriate for slit lamp.
14. Joy stick movement facility for horizontal, vertical, upward and downward adjustment.
15. Light source: 220-240V, 50-60Hz
16. Eye piece: Power range;  $\pm 5D$  or More, IPD – 52-77 mm or more
17. Table: Hydraulic or Motorized.
18. Upgradable to digital photo slit lamp
19. European CE/ US FDA certified / ISO certified
20. Demonstration of the slit lamp before opening the price bid.

**Standard Accessories for each unit of Slit Lamp.**

- |  |            |
|--|------------|
| 1. Chinrest papers   | 05 Packets |
| 2. Dust cover  | 01         |
| 3. Fuses   | 10         |
| 4. Reflecting Mirrors  | 01         |
| 5. Applanation prism   | 02         |
| 6. Co observer tube on beam splitter with focus adjustment and slit rotation facility. |            |

**Terms & Conditions**

1. Warranty 5 years.
2. CAMC 5 years after expiry of 5 years of warranty, including of all required spares and LED

Certified that these specifications are not tailor made and do not pertain to any particular brand

The last date of submission of tender documents may be extended up to 13-11-2018, 12.00 hrs and to open on 13/11/2018 at 12:30 hrs.

sd/

Faculty In-charge  
Procurement Cell  
AIIMS, Patna