



CORRIGENDUM/ADDENDUM

Date: 25/03/2019

With reference to tender PC/2018-19/RC/15133 dated 06/03/2019 tender for Empanelment for Total HIP Replacement (THR) Implant, pre-bid meeting was held on 12-03-2019 at 12:00 PM in conference hall main building AIIMS Patna. The following corrigendum has been issued:-

S N	Tender Specification	Corrigendum
1	Under point C.1.1 – Fully hydroxyapatite coated femoral primary stem (HA coated 155 micron)	Fully or proximally coated hydroxyapatite coated femoral primary stem (HA coated 155 micron)
2.	Metallurgy - Ortron 90	Deleted

3. All the tapers mentioned in the tender to be read as – Stem tapers should be compatible with all the heads quoted by the company.

4. Addendum

- Point E – Dual Mobility cup of all sizes to be quoted by companies.
- Point F – Multihole revision Gription coated cup of all sizes to be quoted by companies.
- Point G – different types of mesh for acetabular wall defect to be quoted by companies.
- Point H – Constrain cups to be quoted by the companies.
- Point I – All sizes Oxinium head to be quoted by the companies.

5. The last date of submission of tender documents has been extended up to 08-04-2019 up-to 17.00 hrs and bids will open on 09/04/2019 at 12:00 noon.

sd/
Faculty In-charge
Procurement Cell
AIIMS Patna