



Dated: 08/02/2018

ADDENDUM

With reference to tender no AIIMS/Pat/Tender/RC/Furniture/Sec-2/17-18/7007 dated 19/01/2018 regarding rate contract for Supply of General Hospital Furniture at AIIMS Patna, a pre-bid meeting was held on 29/01/2018. The committee recommended following addendum to be issued

1. Technical of Phlebotomy chairs in the list of Hospital Furniture. Specification of the item is as follows

Hygienically designed phlebotomy chairs suitable for taking blood in health centres, clinics and hospitals:

- Minimum width 20 inches.
- Framework made of Stainless steel.
- Padded seat, back and armrests.
- Comfortable stain resistant and washable upholstery with good density.
- Adjusting back rest via gas jack, flexible from 95 degrees to 175 degrees.
- Adjusting non slippery leg rest via gas jack, flexible from 0 degrees to 90 degrees.
- Comfortable soft and adjustable head rest.
- Equipped with 2 broad blood test splints, vertically and horizontally adjustable by serrated blocks system lock mechanism.
- Armrest, preferably on left side to be provided with fully laminated storage drawer.
- Light weight and should support user weight of maximum 130kg.
- Made of atleast 40mm deep foam for comfort, antibacterial material used in making
- Available in sober colours: suitable for hospital use.
- Supplied fully assembled – ready for use
- Certification by quality control monitoring body
- Should come with 2 year guarantee

Phlebotomy Chair



(This is only a representative / indicative picture. It does not exactly meet the specifications mentioned below)

sd/
Faculty In-charge
Procurement Cell
AIIMS, Patna \