




CORRIGENDUM

Date: 15/11/2016

With reference to Tender No. AIIMS/Pat/Tender/PC/2015/RadioDiag/CTScan/317, dated 13/10/2016, a pre bid meeting was held on 22/10/2016 at 15:00 Hrs. in Committee Room Main Building AIIMS Patna.

The issue raised by the firm where discussed and the committee recommended for the following corrigendum:

1. Point 3a "Tube Current: minimum range **20-600mA**" may be read as "Tube Current: minimum range **20-800mA**".
2. Point 6d "Single continuous 'spiral-on time' should be minimum 100 seconds or more" should be read as "Single continuous 'spiral-on time' should be minimum 60 seconds or more".
3. Point 10a "Should have two **19" or larger high resolution LED dual-monitor** systems with display matrix of 1024 X 1280 matrix or more, with **two separate keyboards**" should be read as "Should have **19" or larger high resolution LED monitor** with display matrix of 1024 X 1280 matrix or more".
4. Point 11A "A vendor owned dedicated DICOM 3.0 compliant Client Server Architecture based solution should be offered (Intellispace Portal 6/ Dexus-AW server 2/ SyngoVia VA30 or latest). The server should have image storage capacity of at least **10 TB**. A reputed antivirus solution for server and workstations should be in place. The server should be with **4 concurrent users (4 Multimodality Workstations)** facility with workstation memory separate from the console and automatic two way data transfer between the operator console and workstations and server. The workstations will be kept in trauma center and main department reporting rooms. Networking to be done by vendor. All workstations should have following processing capabilities as well as filming and CD and DVD archiving facility independently without the help of main console" should be replaced with "One multimodality workstation should be provided, with automatic two way data transfer, post processing, filming, CD and DVD writing".
5. Point 11A (j)...deleted
6. Point 11D....deleted
7. Point 12f (iv)....deleted
8. Added point 12f (vii).... bone densitometry
9. Added point 12f (viii)... A RIS solution should be offered with minimum 10 licenses. Should be fully compliant with IHE, HL7, and latest IMR standards. Should have registration, billing, scheduling, worklist, reporting, voice recognition and dictation capability, MWL, MPPS.
10. Point 15c....deleted
11. Point 16A "One anaesthesia workstation machine with one multi parameter monitor to ensure smooth completion of cases under general anaesthesia" should be read as "One anaesthesia workstation machine with ventilator and multi parameter monitor to ensure smooth completion of cases under general anaesthesia, suitable for adult, pediatric and neonates"
12. Added point 16A (4).... Specification for ventilator: 1) Should be integrated with the anesthesia system, with bag in bottle ascending type bellows. 2) Should have tidal volume range from 20 ml to 1500 ml. 3) Should be able to set TV, RR and I:E ratio. 4) Should have large display with touch screen and wheel based adjustments. 5) Shall have the following modes: VCV, PCV, SIMV & Pressure Support 6) Should monitor and display integrated Oxygen monitoring, Inspired and expired volumes, PAW, Pressure waveform, Flow waveform and spirometry loops 7) Should have Inspiratory pause, Sigh function, Patient trigger 5 -15 lpm. 8) Should have a max flow of 100 litres per minute in PCV mode. 9) Ventilator should



::2::

- provide all user alarms. 10). Should provide Fresh gas compensation and Compliance compensation. 11). Shall have an active proportional exhalation valve to prevent the potential of over delivery during pressure modes of ventilation. 12). On switching on, the ventilator system check should be able to bypass in the case of an emergency. 13). Flow sensors measurement at the patient end of the circuit is preferred. 14). Should be supplied with necessary reusable and disposable breathing circuits.
13. Added point 16A (5)...etCO2 facility should be present.
14. Point 17 "Suitable double barrel pressure injector with complete accessories & 300 syringes, 1000 tubings and connectors. Mention and freeze the price of the consumables for further purchasing." should be read as "Suitable double barrel pressure injector with complete accessories & 100 pairs of syringes, 200 tubings and connectors. Should have multiphasic and dual flow capability. Injection from console should be possible. Mention and freeze the price of the consumables for further purchasing."
15. Point 18d...deleted
16. Point 18e "One **color laser printer** from reputed manufacturer for printing color images and angiograms. 50 cartridges/toner to be provided" should be read as "One **B&W laser printer-scanner-copier** from reputed manufacturer for printing on paper. 50 cartridges/toner to be provided"
17. *The last date of submission of tender documents is extended up to 01-12-2016 up to 14.00 hrs*

Faculty In-charge
Procurement Cell
AIIMS, Patna