



## CORRIGENDUM

Date: 22/08/2016

With reference to Tender No. AIIMS/Pat/Tender/OphthalmologyOCT/2016/408, dated 3/8/2016, a pre bid meeting was held on 11/082016 at 3:30 pm in conference hall main building AIIMS Patna.

The issue raised by the firms were discussed and the committee recommended following corrigendum issued:-

1. **In technical specification- Point no. 2:** The system should have simultaneous Infrared fundus and OCT imaging **should be read as:** The system should have simultaneous Infrared fundus/red free and OCT imaging.
2. **In technical specification-** The system should have Enhanced depth imaging to view Choroidallayer **should be read as:** The system should have Enhanced depth imaging or swept source technology to view and measure choroidal thickness.
3. *The last date of submission of tender documents is extended up to **08-09-2016 up to 14.00 hrs***

sd/-  
Faculty In-charge  
Procurement Cell  
AIIMS, Patna