



CORRIGENDUM

Date: 22/08/2016

With reference to Tender No. AIIMS/Pat/Tender/OphthalmologyFluorescein/2016/412, dated 3/8/2016, a pre bid meeting was held on 11/082016 at 3:30 pm in conference hall main building AIIMS Patna.

The issue raised by the firms were discussed and the committee recommended following corrigendum issued:-

1. **In technical specification- Point no. 1:** 5 mode multi-modality imaging device allowing, Fluorescein Angiography (FA), Indo Cyanine Green Angiography (ICGA), Fundus Auto Fluorescein (FAF), Infra-Red (IR) and red free imaging **should be read as:** Multi-modality imaging device allowing, Fluorescein Angiography (FA), Indo Cyanine Green Angiography (ICGA), Fundus Auto Fluorescein (FAF), Infra-Red (IR) or red free imaging.
2. **In technical specification-Point no. 6:** Digital camera for colour & FA; At least 5-mega pixels true medical grade camera **should be read as:** Digital camera for colour & FA; At least 5-mega pixels digital camera.
3. *The last date of submission of tender documents is extended up to **08-09-2016 up to 14.00 hrs***

sd/-
Faculty In-charge
Procurement Cell
AIIMS, Patna