



Date: 28/10/2016

Corrigendum/Addendum

With reference to tender no: **AIIMS/Pat/Tender/Bioch/Instrument/599 DATED: 08/10/2016** a pre-bid meeting was held on 18/10/2016 in committee hall, Medical college building, AIIMS Patna.

The issue raised by the firms were discussed and the committee recommended for following corrigendum issued.

For item no.-1 (Deep Freezer -20)

1. Suitable UPS / Servo stabilizer. **Will be read as** “suitable UPS/ Servo stabilizer should be changed with 3 KVA online UPS and suitable Servo stabilizer.”
2. The equipment must be ISO & CE Certified.

**sd/
Faculty In-charge
Procurement Cell
AIIMS, Patna**